

Wavefront aberration measurements and corrections through thick tissue using fluorescent microsphere reference beacons.

Journal: Opt Express

Publication Year: 2010

Authors: Oscar Azucena, Justin Crest, Jian Cao, William Sullivan, Peter Kner, Donald Gavel, Daren Dillon, Scot Olivier, Joel Kubby

PubMed link: 20721137

Funding Grants: AO Wide-Field Microscope

Public Summary:

We present a new method to directly measure and correct the aberrations introduced when imaging through thick biological tissue. A Shack-Hartmann wavefront sensor is used to directly measure the wavefront error induced by a *Drosophila* embryo. The wavefront measurements are taken by seeding the embryo with fluorescent microspheres used as "artificial guide-stars." The wavefront error is corrected in ten millisecond steps by applying the inverse to the wavefront error on a micro-electro-mechanical deformable mirror in the image path of the microscope. The results show that this new approach is capable of improving the Strehl ratio by 2 times on average and as high as 10 times when imaging through 100 microm of tissue. The results also show that the isoplanatic half-width is approximately 19 microm resulting in a corrected field of view 38 microm in diameter around the guide-star.

Scientific Abstract:

We present a new method to directly measure and correct the aberrations introduced when imaging through thick biological tissue. A Shack-Hartmann wavefront sensor is used to directly measure the wavefront error induced by a *Drosophila* embryo. The wavefront measurements are taken by seeding the embryo with fluorescent microspheres used as "artificial guide-stars." The wavefront error is corrected in ten millisecond steps by applying the inverse to the wavefront error on a micro-electro-mechanical deformable mirror in the image path of the microscope. The results show that this new approach is capable of improving the Strehl ratio by 2 times on average and as high as 10 times when imaging through 100 microm of tissue. The results also show that the isoplanatic half-width is approximately 19 microm resulting in a corrected field of view 38 microm in diameter around the guide-star.

Source URL: <https://www.cirm.ca.gov/about-cirm/publications/wavefront-aberration-measurements-and-corrections-through-thick-tissue-using>